



July 24, 2017

**SUPPLEMENTAL/BID BULLETIN**

**ADDENDUM NO. 3**

**SUBJECT : SUPPLY AND DELIVERY OF THERMAL IMAGING SCANNER**

*Reference No. PB170515-RM29/GTMD02-17-95134*

**TO ALL CONCERNED:**

Please be informed of the following amendments on the technical specifications and the deadline of Submission of Bids for the above subject requirement:

ISSUES	AGREEMENTS	
<b>1. Amendment on the Technical Specifications</b>	> Amendments on the Technical Specifications:	
	FROM	TO
	THERMAL SENSITIVITY <0.04°C @ 30°C (minimum)	THERMAL SENSITIVITY ≤0.04°C ~ 0.10°C @ 30°C (minimum)
	FIELD OF VIEW / MINIMUM DISTANCE 25° X 19°/0.25 METERS (minimum)	FIELD OF VIEW / MINIMUM DISTANCE 25° X 19°/0.10 METERS ~ 0.25 METERS (minimum)
	FRAME RATE = 30 HZ	FRAME RATE = 30 HZ ~ 60HZ (minimum)
PERIODIC IMAGE STORAGE: 7 SECONDS TO 24 HOURS (IR) / 14 SECONDS TO 24 HOURS (IR AND VISUAL)	PERIODIC IMAGE STORAGE: 7 SECONDS TO 24 HOURS (IR) / 14 SECONDS ~ 24 HOURS (IR AND VISUAL) or AUTOCAPTURE FUNCTION	

Please be guided accordingly.

**GENEROSO M. SENAL**

Chairperson  
Bids and Awards Committee